

Art.-No.	Width	Length	Height
CW 12.14.12	12 mm	14 mm	12 mm
CW 12.14.14	12 mm	14 mm	14 mm
CW 12.14.16	12 mm	14 mm	16 mm
CW 12.14.18	12 mm	14 mm	18 mm
CW 12.14.20	12 mm	14 mm	20 mm
CW 12.14.22	12 mm	14 mm	22 mm
CW 12.14.24	12 mm	14 mm	24 mm
CW 12.14.26	12 mm	14 mm	26 mm

These vertebral body replacement implants can be combined with all CERVICUS^{Titan} cages (PINA) with Art.-No. CK12.14.04 – CK12.14.09.

Art.-No.	Width	Length	Height
CW 14.16.12	14 mm	16 mm	12 mm
CW 14.16.14	14 mm	16 mm	14 mm
CW 14.16.16	14 mm	16 mm	16 mm
CW 14.16.18	14 mm	16 mm	18 mm
CW 14.16.20	14 mm	16 mm	20 mm
CW 14.16.22	14 mm	16 mm	22 mm
CW 14.16.24	14 mm	16 mm	24 mm
CW 14.16.26	14 mm	16 mm	26 mm

These vertebral body replacement implants can be combined with all CERVICUS^{Titan} Cages (PINA) with Art.-No. CK14.16.04 – CK14.16.09.

PINA[®]

Your Swiss Option

VOLUTUS^{Titan}



PINA Medizintechnik Vertriebs AG
 Ebringer Str. 161 CH 8240 Thayngen
 Phone +41 52 672 4042 www.pina-med.ch

Description

The VOLUTUS^{Titan} cervical vertebral body replacements from PINA® Medizintechnik Vertriebs AG are implants intended for implantation between the end plates of the individual vertebrae. Their shape is adapted to the morphology of the vertebral bodies being treated and the prescribed surgical techniques. The body replacements are available in various designs and sizes. VOLUTUS^{Titan} cervical vertebral body replacement is intended for use below C3 to C7.



Product characteristic

VOLUTUS^{Titan} cervical vertebral body replacements are made from the Titanium alloy ISO 5832-3. A large contact area allows a maximal bone contact and reduces the risk of subsidence. Through the existing openings, a filling with autologous bone or bone substitute material is possible. A special feature is the modular design. This modular system covers all intervertebral heights and all vertebral body heights of the cervical spine with the exception of C1 and C2 variable and allows the correct reconstruction of the lordosis.



Indications

- Ventral corpectomy with cervical myelopathy
- Tumours in the cervical vertebral bodies
- Comminuted fractures of cervical vertebral bodies

Width	Length	Height
12 mm	14 mm	12 mm
12 mm	14 mm	14 mm
12 mm	14 mm	16 mm
12 mm	14 mm	18 mm
12 mm	14 mm	20 mm
12 mm	14 mm	22 mm
12 mm	14 mm	24 mm
12 mm	14 mm	26 mm

These vertebral body replacement implants can be combined with all CERVICUS^{Titan} cages (PINA) with Art.-No. CK12.14.04 – CK12.14.09.

Width	Length	Height
14 mm	16 mm	12 mm
14 mm	16 mm	14 mm
14 mm	16 mm	16 mm
14 mm	16 mm	18 mm
14 mm	16 mm	20 mm
14 mm	16 mm	22 mm
14 mm	16 mm	24 mm
14 mm	16 mm	26 mm

These vertebral body replacement implants can be combined with all CERVICUS^{Titan} cages (PINA) with Art.-No. CK14.16.04 – CK14.16.09.

Instruments VOLUTUS^{Titan}

